

Rheumatoid Arthritis

ichromaTM RF IgM

INTENDED USE

ichromaTM RF IgM is a fluorescence immunoassay (FIA) for the quantitative determination of RF IgM in human whole blood/serum/ plasma. It is useful as an aid in management and monitoring of rheumatoid arthritis.

For *in vitro* diagnostic use only.

INTRODUCTION

Rheumatoid arthritis (RA) is the most common chronic autoimmune arthritis worldwide, leading to disability and substantial economic costs. It is a chronic and systemic inflammatory disorder that may affect many tissues and organs, but principally attacks synovial joints. About 1% of the world's population is afflicted by rheumatoid arthritis, women three times more often than men. Onset is most frequent between the ages of 40 and 50, but people of any age can be affected. It can be a disabling and painful condition, which can lead to substantial loss of functioning and mobility if not adequately treated.

PRINCIPLE

The test uses a sandwich immunodetection method.

The detector antibodies in buffer bind to antigens in the sample, forming antigen-antibody complexes, and migrate onto nitrocellulose matrix to be captured by the other immobilized-antibody on a test strip.

More antigens in the sample will form more antigen-antibody complexes which lead to stronger fluorescence signal by detector antibodies, which is processed by the instrument for ichromaTM tests to show RF IgM concentration in the sample.

COMPONENTS

ichromaTM RF IgM consists of 'cartridges', 'detector tubes' and 'detector diluent'.

- The cartridge contains the membrane called a test strip, which has Purified Human IgG at the test line, and streptavidin at the control line. All cartridges are individually sealed in an aluminum foil pouch containing a desiccant, and they are further packaged in a box.
- The detector tube has a granule containing anti-IgM fluorescence conjugate and BSA-Biotin fluorescence conjugate in phosphate buffered saline (PBS). All detector tubes are packed in a pouch.
- The detector diluent contains tween20 as a surfactant, and sodium azide as a preservative in potassium phosphate buffer, and it is dispensed in a vial. The detector diluent is packed in a box.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Follow the instructions and procedures described in this 'Instructions for use'.
- Use only fresh samples and avoid direct sunlight.

- Lot numbers of all the test components (cartridge, detector tube, detector diluent and ID chip) must match each other.
- Do not interchange the test components between different lots or use the test components after the expiration date, either of which might yield incorrect test result(s).
- Do not reuse cartridges or detector tubes. A cartridge should be used for testing one sample only. A detector tube should be used for processing of one sample only.
- The cartridge should remain sealed in its original pouch until just before use. Do not use cartridge, if pouch is damaged or has already been opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with local regulations. Sample with severe hemolysis and/or hyperlipidemia must not be used.
- If test components and/or sample are stored in refrigerator, then allow cartridge, detector tube, detector diluent and sample to be at room temperature for approximately 30 minutes before use.
- The instrument for ichromaTM tests may generate slight vibration during use.
- Used cartridges, detector tubes, detector diluent and pipette tips should be handled carefully and discarded by an appropriate method in accordance with relevant local regulations.
- The detector diluent contain sodium azide (NaN₃), and they may cause certain health issues like convulsions, low blood pressure and heart rate, loss of consciousness, lung injury and respiratory failure. Avoid contact with skin, eyes, and clothing. In case of contact, rinse immediately with running water.
- No Biotin interference was observed in **ichromaTM RF IgM** when biotin concentration in the sample was below 300 ng/mL. If a patient has been taking biotin at dosage of more than 0.03 mg a day, it is recommended to test again 24 hours after discontinuation of biotin intake.
- **ichromaTM RF IgM** will provide accurate and reliable results subject to the below conditions.
 - **ichromaTM RF IgM** should be used only in conjunction with the instrument for ichromaTM tests.
 - Have to use recommended anticoagulant.

Recommended anticoagulants

Na₂ EDTA, K₂ EDTA,
Sodium citrate, Lithium heparin

STORAGE AND STABILITY

Storage condition			
Component	Storage Temperature	Shelf life	Note
Cartridge	2 - 30°C	20 months	Disposable
Detector tube	2 - 30°C	20 months	Unopened
Detector	2 - 30°C	20 months	Unopened
diluent	2 - 30°C	3 months	Opened

- After the cartridge pouch is opened, the test should be performed immediately.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antibodies.
- The test may yield false negative result(s) due to the non-responsiveness of the antigens to the antibodies which is the most common if the epitope is masked by some unknown components, so therefore not being able to be detected or captured by the antibodies. The instability or degradation of the antigens with time and/or temperature may also cause false negative result as it makes antigens unrecognizable by the antibodies.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician in conjunction with clinical symptoms and other relevant test results.

MATERIALS SUPPLIED

REF CFPC-39

Components of **ichroma™ RF IgM**

- | | |
|------------------------|----|
| ■ Cartridge box: | |
| - Cartridge | 25 |
| - Detector tube | 25 |
| - Detector diluent | 1 |
| - ID chip | 1 |
| - Instructions for use | 1 |

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately with **ichroma™ RF IgM**.

Please contact our sales division for more information.

- | | |
|----------------------------------|---------------------|
| ■ Instrument for ichroma™ tests | |
| - ichroma™ Reader | REF FR203 |
| - ichroma™ II | REF FPRR021 |
| - ichroma™ III | REF FPRR037 |
| - ichroma™ M3 | REF FPRR035 |
| ■ Printer | REF FPRR007 |
| ■ Boditech RF IgM Control | REF CFPO-103 |

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ RF IgM** is human whole blood/serum /plasma.

- It is recommended to test the sample within 24 hours after collection when collected sample is stored at room temperature.
- The samples (serum, plasma) should be separated from the clot by centrifugation within 3 hours after the collection of whole blood.
- The samples (serum, plasma) may be stored for a week at 2-8 °C prior to being tested. If testing will be delayed more than a week, samples (serum, plasma) should be frozen at -20 °C.
- The samples (serum, plasma) stored frozen at -20 °C for 2 months showed no performance difference.
- However, the whole blood sample should not be kept in a freezer in any case.

- As a repeated freeze thaw cycle may affect the test result, do not refreeze previously frozen samples.

TEST SETUP

- Check the contents of **ichroma™ RF IgM**: Sealed cartridges, detector tubes, a detector diluent, an ID chip, and an instructions for use.
- Ensure that the lot number of the cartridges matches that of the detector tube, the detector diluent as well as an ID chip.
- If the sealed cartridge, the detector tube and the detector diluent have been stored in a refrigerator, place them on a clean and flat surface at room temperature for at least 30 minutes before testing.
- Turn on the instrument for ichroma™ tests.
- Insert the ID chip into the 'ID chip port'.
- ※ Please refer to the instrument for ichroma™ tests operation manual for complete information and operating instructions.

TEST PROCEDURE

► **ichroma™ Reader, ichroma™ II, ichroma™ M3** **Multi test mode**

- 1) Select the sample type (whole blood or serum/ plasma) on the screen.
- 2) Take 150 µL of detector diluent using a pipette and dispense it to the detector tube containing a granule. When the granule form is completely dissolved in the tube, it becomes detection buffer.
(The detection buffer must be used immediately. Do not exceed 30 seconds.)
- 3) Take 10 µL (Human whole blood) or 5 µL (Human serum/plasma/control) of sample using a pipette and dispense it to the detector tube.
- 4) Close the lid of the detector tube and mix the sample thoroughly by shaking it about 10 times.
(The sample mixture must be used immediately. Do not exceed 30 seconds.)
- 5) Take 75 µL of the sample mixture and dispense it into the sample well of the cartridge.
- 6) Leave the cartridge at room temperature for 5 minutes.
⚠ Scan the sample-loaded cartridge immediately when the incubation time is over. If not, it will cause inaccurate test result.
- 7) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 8) Press the 'Select' or tap the 'Start' button on the instrument for ichroma™ tests to start the scanning process.
(ichroma™ M3 will start the test automatically after inserting.)
- 9) The instrument for ichroma™ tests will start scanning the sample-loaded cartridge immediately.
- 10) Read the test result on the display screen of the instrument for ichroma™ tests.

Single test mode

- 1) The test procedure is same with the 'Multi test mode 1) – 4)'.
- 2) Insert the sample-loaded cartridge into the holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 3) Press the 'Select' or tap the 'Start' button on the instrument for ichroma™ tests.
(ichroma™ M3 will start the test automatically after inserting.)
- 4) The cartridge goes inside the instrument for ichroma™ tests and will automatically start scanning the sample-loaded cartridge after 5 minutes.
- 5) Read the test result on the display screen of the instrument for ichroma™ tests.

▶ ichroma™ III

- 1) The test procedure is same with the 'Single test mode'.

INTERPRETATION OF TEST RESULT

- The instrument for ichroma™ tests calculates the test result automatically and displays RF IgM concentration of the test sample in terms of IU/mL.
- Cut-off value : 15 IU/mL
- Working range : 8-200 IU/mL

QUALITY CONTROL

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are provided on demand with **ichroma™ RF IgM**. For more information regarding obtaining the control materials, contact Boditech Med Inc.'s Sales Division for assistance.
(Please refer to the instructions for use of control material.)

PERFORMANCE CHARACTERISTICS

- **Analytical sensitivity:**
 - Limit of Blank (LoB) 2.63 IU/mL
 - Limit of Detection (LoD) 3.67 IU/mL
 - Limit of Quantitation (LoQ) 7.78 IU/mL
- **Analytical specificity:**
 - Cross-reactivity
Cross-reactivity test is not considered because RF is non-specific for non-rheumatic and healthy persons.
 - Interference
Interferents listed in the following table were added to the test at the concentration mentioned below. **ichroma™ RF IgM** test results did not show any significant interference with these materials.

Interferents	Concentration
Hemoglobin(human)	1000 mg/dL
Bilirubin[unconjugate]	40 mg/dL
triglyceride	2000 mg/dL
Human serum albumin	12 g/dL

Methotrexate	1.36 mg/mL
Sulfasalazine	75 µg/mL

■ Precision

- Single-site study

Repeatability (within-run precision)

within-laboratory precision (Total precision)

Lot to lot precision

3 Lots of **ichroma™ RF IgM** were tested for 20 days. Each standard material was tested 2 times per day. For each test, each material was duplicated.

Single-site study							
RF IgM [IU/mL]	Repeatability		within-laboratory precision		Lot to lot precision		
	AVG	CV	AVG	CV	AVG	CV	
	[IU/mL]	(%)	[IU/mL]	(%)	[IU/mL]	(%)	
14	14.31	7.52	14.39	7.53	14.20	7.50	
54	53.39	6.14	53.86	5.92	54.03	6.31	
109	107.12	6.08	108.27	6.08	109.14	6.18	

- Multi-site study

Reproducibility

1 Lot of **ichroma™ RF IgM** was tested for 5 days in 3 different sites (1 person per 1 site, 1 instrument per 1 site). Each standard material was tested 1 time per and 5 replicates per day.

Multi-site study			
RF IgM [IU/mL]	Reproducibility		
	AVG [IU/mL]	CV (%)	
14	13.91	6.29	
54	54.72	7.26	
109	109.81	7.89	


■ Accuracy

The accuracy was confirmed by testing with 3 different lots of **ichroma™ RF IgM**. The tests were repeated 10 times at each concentration of the control standard.

RF IgM [IU/mL]	Lot 1	Lot 2	Lot 3	Mean	Recovery (%)
14	13.96	14.62	14.52	14.36	102.6
27	26.49	27.27	27.51	27.09	100.3
54	55.39	54.89	54.45	54.91	101.7
82	82.81	78.70	85.30	82.27	100.9
109	108.34	107.17	110.10	108.54	99.6
180	187.31	173.17	181.14	180.54	100.3

■ **Comparability**

RF IgM concentrations of 100 clinical samples were quantified independently with **ichroma™ RF IgM (ichroma™ II)** and **comparator A** as per prescribed test procedures. Test results were compared, and their comparability was investigated with linear regression and correlation coefficient (R). The regression equation and correlation coefficient are as follows.



REFERENCES

1. Franziska Renge. Immediate determination of ACPA and rheumatoid factor - a novel point of care test for detection of anti-MCV antibodies and rheumatoid factor using a lateral-flow immunoassay. 2010. Arthritis Res Ther. 12:R120.
2. Hiroshi Oka, Shunsei Hirohata. Quantitative determination of IgM-rheumatoid factor by enzyme immunoassay - Standardization using a serum from a rheumatoid arthritis patient. 1990. 188:147-159.
3. Williams RC Jr. Autoimmune mechanisms involved in the pathogenesis of rheumatoid arthritis. 1996. Adv Dent Res. 10(1):47-51.
4. Gioud-Paquet M, Auvinet M, Raffin T, Girard P. IgM rheumatoid factor (RF), IgA RF, IgE RF, and IgG RF detected by ELISA in rheumatoid arthritis. 1987. Ann Rheum Dis. 46(1):65-71.
5. Kawashiri SY, Kawakami A, Ueki Y, Imazato T, Iwamoto N, Fujikawa K, Aramaki T, Tamai M, Nakamura H, Origuchi T, Ida H, Eguchi K. Decrement of serum cartilage oligomeric matrix protein (COMP) in rheumatoid arthritis (RA) patients achieving remission after 6 months of etanercept treatment: comparison with CRP, IgM-RF, MMP-3 and anti-CCP Ab. 2010. Joint Bone Spine. 77(5):418-20.
6. Aletaha D, Neogi T, Silman AJ, et al. 2010 Rheumatoid arthritis classification criteria: an American College of Rheumatology/European League Against Rheumatism collaborative initiative. Arthritis Rheum. 2010 Sep;62(9):2582-91.
7. Lipsky PE. Rheumatoid arthritis. In : Fauci, Braunwald, Kasper, Hauser, Longo, Jameson, Loscalzo, eds. Harrison's 2008, 2083-92.
8. Christopher V, Tehlirian, Joan M, et al. Rheumatoid Arthritis, A clinical and laboratory manifestations In : Klippel JH, Stone JH, Crofford LJ, White PH. eds. Primer on the rheumatic disease, 13th ed. New York : Springer, 2008, 114-121.
9. Scott DL, Symmons DP, Coulton BL, Popert AJ. Long-term outcome of treating rheumatoid arthritis: results after 20 years. Lancet 1987;1:1108-11.
10. Tobón GJ, Youinou P, Saraux A. The environment, geo-epidemiology, and autoimmune disease: rheumatoid arthritis. Autoimmun Rev. 2010 Mar;9(5):A288-92.
11. Alamanos Y, Drosos AA. Epidemiology of primary Sjögren's syndrome in north-west Greece, 1982-2003. Rheumatology (Oxford). 2006 Feb;45(2):187-91.
12. McInnes IB, Schett G. The Pathogenesis of Rheumatoid Arthritis. N Engl J Med. 2011 Dec 8;365(23):2205-19.

Note: Please refer to the table below to identify various symbols

	Sufficient for <n> tests
	Read instruction for use
	Use by Date
	Batch code
	Catalog number
	Caution
	Manufacturer
	Authorized representative of the European Community
	In vitro diagnostic medical device
	Temperature limit
	Do not reuse
	This product fulfills the requirements of the Directive 98/79/EC on in vitro diagnostic medical devices

For technical assistance, please contact:

Boditech Med Inc.'s Technical Sales

Tel: +(82) -33-243-1400

E-mail: TS@boditech.co.kr



Boditech Med Inc.

43, Geodudanji 1-gil, Dongnae-myeon, Chuncheon-si, Gang-won-do, 24398, Republic of Korea

Tel: +(82) -33-243-1400

Fax: +(82) -33-243-9373

www.boditech.co.kr



Obelis s.a

Bd. Général Wahis 53, 1030 Brussels, Belgium

Tel: +(32) -2-732-59-54

Fax: +(32) -2-732-60-03

E-Mail: mail@obelis.net

