



ichroma™ Ferritin

INTENDED USE

ichroma™ Ferritin is a fluorescence immunoassay (FIA) for the quantitative determination of Ferritin in human serum/plasma. It is useful as an aid for quantifying human ferritin.
For *in vitro* diagnostic use only.

INTRODUCTION

Ferritin, a major iron storage protein, is essential to iron homeostasis and is involved in a wide range of physiologic and pathologic processes. Ferritin makes iron available for critical cellular processes while protecting lipids, DNA, and proteins from the potentially toxic effects of iron. In clinical medicine, ferritin is predominantly utilized as a marker of total body iron stores. In cases of iron deficiency and overload, serum ferritin serves a critical role in both diagnosis and management. It is clear that low ferritin values less than reference range are usually representative of body iron deficiency. Recent study suggests that ferritin provides a more sensitive, specific and reliable measurement for determining iron deficiency at an early stage. On other hand, patients with ferritin levels that are higher than the reference range may be indicative of conditions such as iron overload, infections, inflammations, collagen diseases, hepatic diseases, neoplastic disease and chronic renal failure.

PRINCIPLE

The test uses a sandwich immunodetection method.
The detector antibodies in buffer bind to antigens in the sample, forming antigen-antibody complexes, and migrate onto nitrocellulose matrix to be captured by the other immobilized antibodies on a test strip.
More antigens in the sample will form more antigen-antibody complexes which lead to stronger fluorescence signal by detector antibodies, which is processed by the instrument for ichroma™ tests to show ferritin concentration in the sample.

COMPONENTS

ichroma™ Ferritin consists of 'cartridges', 'detector tubes' and 'detector diluent'.

- The cartridge contains the membrane called a test strip which has anti-Ferritin at the test line, and streptavidin at the control line. All cartridges are individually sealed in an aluminum foil pouch containing a desiccant, and they are further packaged in a box.
- The detector tube has a granule containing anti-Ferritin-fluorescence conjugate, Biotin-BSA-fluorescence conjugate in phosphate buffered saline (PBS). All detector tubes are packed in a pouch.
- The detector diluent contains sodium azide as a preservative in phosphate buffered saline (PBS), and it is pre-dispensed in a vial. The detector diluent is packed in a box.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Follow instructions and procedures described in this 'Instructions for use'.
- Use only fresh samples and avoid direct sunlight.
- Lot numbers of all the test components (cartridge, detector tube, detector diluent and ID chip) must match each other.
- Do not interchange the test components between different lots or use the test components after the expiration date, either of which might yield incorrect test result(s).
- Do not reuse cartridges or detector tubes. A cartridge should be used for testing one sample only. A detector tube should be used for processing of one sample only.
- The cartridge should remain sealed in its original pouch until just before use. Do not use the cartridge, if pouch is damaged or has already been opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with local regulations. Sample with severe hemolysis and/or hyperlipidemia must not be used.
- If test components and/or sample are stored in refrigerator, then allow cartridge, detector tube, detector diluent and sample to be at room temperature for approximately 30 minutes before use.
- The instrument for ichroma™ tests may generate slight vibration during use.
- Used cartridges, detector tubes, detector diluent and pipette tips should be handled carefully and discarded by an appropriate method in accordance with relevant local regulations.
- The detector diluent contains sodium azide (NaN₃), and it may cause certain health issues like convulsions, low blood pressure, low heart rate, loss of consciousness, lung injury and respiratory failure. Avoid contact with skin, eyes, and clothing. In case of contact, rinse immediately with running water.
- No Biotin interference was observed in **ichroma™ Ferritin** when biotin concentration in the sample was below 20 ng/mL. If a patient has been taking biotin at dosage of more than 0.03 mg a day, it is recommended to test again 24 hours after discontinuation of biotin intake.
- **ichroma™ Ferritin** will provide accurate and reliable results subject to the below conditions.
 - **ichroma™ Ferritin** should be used only in conjunction with the instrument for ichroma™ tests.
 - Have to use recommended anticoagulant.

Recommended anticoagulant

K₂ EDTA, Sodium citrate, Sodium Heparin

STORAGE AND STABILITY

Storage condition			
Component	Storage Temperature	Shelf life	Note
Cartridge	2 – 30°C	20 months	Disposable
Detector tube	2 – 30°C	20 months	Disposable
Detector diluent	2 – 30°C	20 months	Unopened
		20 months	Opened

- After the cartridge pouch is opened, the test should be performed immediately.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antibodies.
- The test may yield false negative result(s) due to the non-responsiveness of the antigens to the antibodies which is the most common if the epitope is masked by some unknown components, so therefore not being able to be detected or captured by the antibodies. The instability or degradation of the antigens with time and/or temperature may also cause the false negative result as it makes the antigens unrecognizable by the antibodies.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician in conjunction with clinical symptoms and other relevant test results.

MATERIALS SUPPLIED

REF CFPC-32

Components of **ichroma™ Ferritin**

- Cartridge box:
 - Cartridge 25
 - Detector tube 25
 - Detector diluent 1
 - ID chip 1
 - Instructions for use 1

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately with **ichroma™ Ferritin**.

Please contact our sales division for more information.

- Instrument for **ichroma™** tests

- **ichroma™ Reader**
- **ichroma™ II**
- **ichroma™ III**
- **ichroma™ M3**

REF FR203
REF FPRR021
REF FPRR037
REF FPRR035
REF FPRR007
REF FPRR009
REF CFPO-99

- **Printer**
- **i-Chamber**
- **Boditech Ferritin Control**

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ Ferritin** is human serum/plasma.

- It is recommended to test the sample within 24 hours after collection when collected sample is stored at room temperature.
- The samples (serum, plasma) should be separated from the clot by centrifugation within 3 hours after the collection of whole blood.
- The samples (serum, plasma) may be stored for a month at 2 – 8°C prior to being tested. If testing will be delayed more than a month, samples (serum, plasma) should be frozen at -20°C.
- The samples (serum, plasma) stored frozen at -20°C for 3 months showed no performance difference.

- As a repeated freeze-thaw cycle may affect the test result, do not refreeze previously frozen samples.

TEST SETUP

- Check the contents of **ichroma™ Ferritin**: Sealed cartridges, detector tubes, a detector diluent, an ID chip and an instructions for use.
- Ensure that the lot number of the cartridge matches that of the detector tube, the detector diluent as well as the ID chip.
- If the sealed cartridge, the detector tube and the detector diluent have been stored in a refrigerator, place them on a clean and flat surface at room temperature for at least 30 minutes before testing.
- Turn on the instrument for **ichroma™** tests.
- Insert the ID chip into the 'ID chip port'.
- ※ Please refer to the instrument for **ichroma™** tests operation manual for complete information and operating instructions.

CAUTION

- To minimize erroneous test results, we suggest that the ambient temperature of the cartridge should be 25°C during the reaction time after loading sample mixture to the cartridge.
- To maintain the ambient temperature to 25°C, you can use various devices such as an i-Chamber or an incubator and so on.

TEST PROCEDURE

► **ichroma™ Reader, ichroma™ II**

- 1) Select the sample type (whole blood or serum/plasma) on the screen.
 - ※ If using control as the sample, select the 'serum/plasma'.
- 2) Insert the cartridge approximately 1/3 into the slot of the i-Chamber.
- 3) Take 150 µL of detector diluent using a pipette and dispense it to the detector tube containing a granule. When the granule form is completely dissolved in the tube, it becomes detection buffer.
(The detection buffer must be used immediately. Do not exceed 30 seconds.)
- 4) Take 30 µL of sample (serum/plasma/control) using a pipette and dispense it to the detector tube.
- 5) Close the lid of the detector tube and mix the sample thoroughly by shaking it about 10 times.
(The sample mixture must be used immediately. Do not exceed 30 seconds.)
- 6) Take 75 µL of the sample mixture and dispense it into the sample well of the cartridge.
- 7) Wait until the sample mixture flows into the windows, which may take about 10 seconds.
- 8) Insert the sample-loaded cartridge into the slot of the i-Chamber (25°C), an incubator, and leave it for 10 minutes.

⚠ Scan the sample-loaded cartridge immediately when the incubation time is over. If not, it will cause inaccurate test result.

- 9) To scan the sample-loaded cartridge, insert it into the cartridge holder of the Instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 10) Press 'Select' or tap the 'Start' button on the instrument for ichroma™ tests to start the scanning process.
- 11) The instrument for ichroma™ tests will start scanning the sample-loaded cartridge immediately.
- 12) Read the test result on the display screen of the instrument for ichroma™ tests.

► ichroma™ M3

- 1) Select the sample type (whole blood or serum/etc.) on the screen. Press the key button to change the desired sample type.
※ If using control as the sample, select the 'serum/etc.'.
- 2) Insert the cartridge approximately 1/3 into the slot of the ichroma™ M3.
Before pushing the cartridge all the way in, ensure it is inserted in the correct orientation. An arrow is marked on the cartridge especially for this purpose.
- 3) The test procedure is same with the 'ichroma™ II test procedure 3) ~ 6)'.
- 4) Wait until the sample mixture flows into the window.

▲ Insert the cartridge exactly 30 seconds after applying the sample. Otherwise test results may be inaccurate.

- 5) Insert the sample-loaded cartridge fully into the slot of the ichroma™ M3.
- 6) After inserting the cartridge, the result will be automatically displayed after 10 minutes.
- 7) Read the test result on the display screen of the instrument for ichroma™ tests.

► ichroma™ III

- 1) Set the temperature to 25°C by selecting test cartridge type on the screen.
- 2) Select the sample type (whole blood or serum/plasma) on the screen.
※ If using control as the sample, select the 'serum/plasma'.
- 3) The test procedure is same with the 'ichroma™ II test procedure 3) ~ 6)'.
- 4) Insert the sample-loaded cartridge into the holder of ichroma™ III. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 5) Tap 'Start' button on ichroma™ III to start the scanning process.
- 6) The cartridge goes inside and ichroma™ III will automatically start scanning the sample-loaded cartridge after 10 minutes.
- 7) Read the test result on the display screen of the ichroma™ III.

INTERPRETATION OF TEST RESULT

- The instrument for ichroma™ tests calculates the test result automatically and displays Ferritin concentration of the test sample in terms of ng/mL.
- Reference range
 - Women 20 – 250 ng/mL
 - Men 30 – 350 ng/mL
- Working range: 10 – 1,000 ng/mL

QUALITY CONTROL

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are provided on demand with **ichroma™ Ferritin**. For more information regarding obtaining the control materials, contact [Boditech Med Inc.'s Sales Division for assistance](#).
(Please refer to the instructions for use of control material.)

PERFORMANCE CHARACTERISTICS

- **Analytical Sensitivity**
 - Limit of Blank (LoB) 1.56 ng/mL
 - Limit of Detection (LoD) 3.31 ng/mL
 - Limit of Quantitation (LoQ) 10.00 ng/mL

- **Analytical Specificity**
 - Cross reactivity
Biomolecules listed in the following table were added to the test sample(s) at concentrations much higher than their normal physiological levels in the blood. **ichroma™ Ferritin** test results did not show any significant cross-reactivity with these biomolecules.

Cross-reactants	Concentration
Human Transferrin	100 mg/dL
Ferric Chloride	100 mg/dL
Human Serum Albumin	10 g/dL

- Interference
Interferents in the following table were added to the test sample at the concentration mentioned below. **ichroma™ Ferritin** test results did not show any significant interference with these materials.

Interferents	Concentration
Bilirubin(conjugated)	40 mg/dL
Bilirubin(unconjugated)	40 mg/dL
Triglyceride, Total	1,500 mg/dL
Hemoglobin	1,000 mg/dL

■ Precision

3 Lots of **ichroma™ Ferritin** were tested for 21 days (7 days per 1 Lot at 1 site by one reader). Each standard material was tested 2 times per day. For each test, each material was duplicated.

- Repeatability (within-run precision)
Repeatability of **ichroma™ Ferritin** was evaluated with results of 1 Lot.
- Total precision (within-laboratory precision)
Total precision (within-run, between-run, between-day) of **ichroma™ Ferritin** was evaluated with results of 1 Lot.
- Lot to lot precision
Lot to lot precision of **ichroma™ Ferritin** was evaluated with results of 3 Lots.
- Between site
Three persons tested **ichroma™ Ferritin** at 3 different sites, 10 times at each concentration of standard materials.
- Between person
3 persons tested **ichroma™ Ferritin**, 10 times at each concentration of standard materials.
- Between reader
3 different persons tested same lot of **ichroma™ Ferritin** with 3 different readers, 10 times at each concentration of the control standard.

Ferritin [ng/mL]	Repeatability (within-run precision)		Total precision (within-laboratory precision)	
	AVG	CV(%)	AVG	CV(%)
25	25.44	5.6	25.13	5.8
100	99.57	5.6	99.20	6.1
500	501.86	5.3	504.62	5.9
Ferritin [ng/mL]	Lot to lot precision		Between site	
	AVG	CV(%)	AVG	CV(%)
25	24.85	6.2	24.69	6.4
100	100.20	6.4	98.60	5.5
500	504.29	6.0	483.65	6.3
Ferritin [ng/mL]	Between person		Between reader	
	AVG	CV(%)	AVG	CV(%)
25	24.72	5.7	25.02	6.6
100	100.09	5.5	99.44	4.6
500	499.46	5.7	497.02	5.7

■ Accuracy

The accuracy was confirmed by testing with 3 different lots of **ichroma™ Ferritin**. The tests were repeated 10 times at each concentration of the control standard.


Ferritin [ng/mL]	Lot 1	Lot 2	Lot 3	AVG [ng/mL]	Recovery (%)
12.5	12.34	12.18	12.32	12.28	98.25
25	25.52	26.04	25.45	25.67	102.68
100	97.56	98.08	99.84	98.49	98.49
500	489.45	481.61	489.11	486.72	97.34
1,000	971.99	945.99	956.41	958.13	95.81

■ Comparability

Ferritin concentrations of 100 clinical samples were quantified independently with **ichroma™ Ferritin (ichroma™ II)** and **comparator A** as per prescribed test procedures. Test results were compared, and their comparability was investigated with linear regression and correlation coefficient (R). The regression equation and correlation coefficient are as follows.

* Solid line: regression line of **ichroma™ Ferritin**

* Dash line: $y = x$



REFERENCES

- Garcia-Casal MN, Peña-Rosas JP, Urrechaga E, Escanero JF, Huo J, Martinez RX, Lopez-Perez L. Performance and comparability of laboratory methods for measuring ferritin concentrations in human serum or plasma: A systematic review and meta-analysis. PLoS One. 2018 May 3;13(5):e0196576.
- K Morikawa, F Oseko, S Morikawa. A Role for Ferritin in Hematopoiesis and the Immune System. Leukemia and Lymphoma, Vol. 18, pp. 429-433.
- A Hamwi, G Endler, K Rubi, O Wagner, AT Endler. Reference Values for a Heterogeneous Ferritin Assay and Traceability to the 3rd International Recombinant Standard for Ferritin (NIBSC Code 94/572). Clin Chem Lab Med 2002; 40(4):365-370
- W Wang, MA Knovich, LG Coffman, FM Torti, SV Torti. Serum ferritin: Past, present and future. Biochimica et Biophysica Acta 1800 (2010) 760-769
- PM Harrison, P Arosio. The ferritins: molecular properties, iron storage function and cellular regulation. Biochimica et Biophysica Acta 1275 (1996) 161-203

Note: Please refer to the table below to identify various symbols.

	Sufficient for <n> tests
	Read instruction for use
	Use by Date
	Batch code
	Catalog number
	Caution
	Manufacturer
	Authorized representative of the European Community
	<i>In vitro</i> diagnostic medical device
	Temperature limit
	Do not reuse
	This product fulfills the requirements of the Directive 98/79/EC on <i>in vitro</i> diagnostic medical devices

For technical assistance, please contact:

Boditech Med Inc.'s Technical Sales

Tel: +(82) -33-243-1400

E-mail: TS@boditech.co.kr



Boditech Med Inc.

43, Geodudanji 1-gil, Dongnae-myeon, Chuncheon-si,
Gang-won-do, 24398, Republic of Korea
Tel: +(82) -33-243-1400
Fax: +(82) -33-243-9373
www.boditech.co.kr



Obelis s.a

Bd. Général Wahis 53, 1030 Brussels, Belgium
Tel: +(32) -2-732-59-54
Fax: +(32) -2-732-60-03
E-Mail: mail@obelis.net

