



ichroma™ Tn-I Plus

INTENDED USE

ichroma™ Tn-I Plus is a fluorescence immunoassay (FIA) for the quantitative determination of cardiac Tn-I (troponin-I) in human whole blood/serum/plasma. It is useful as an aid in management and monitoring of acute myocardial infarction (AMI).

For *in vitro* diagnostic use only.

INTRODUCTION

Cardiac troponins are currently the most sensitive and specific biochemical markers of myocardial necrosis. There are three types of troponin in heart muscle fibers. Those are troponin-C, troponin-I, and troponin-T. Together they contribute to make cardiac muscle fibers contract. The clinical measurement of serum Tn-I has become an important tool in the diagnosis of acute myocardial infarction. Serum Tn-I is a more reliable than creatine kinase as a prognostic marker in people with ischemic chest pain. National and international scientific organizations have suggested the use of troponins, Tn-I and Tn-T, when implementing new diagnostic strategies in patients with acute coronary syndrome.

PRINCIPLE

The test uses a sandwich immunodetection method.

The detector antibodies in buffer bind to antigens in the sample, forming antigen-antibody complexes, and migrate onto nitrocellulose matrix to be captured by the other immobilized streptavidin on a test strip.

More antigens in the sample will form more antigen-antibody complexes which lead to stronger fluorescence signal by detector antibodies, which is processed by the instrument for ichroma™ tests to show Tn-I concentration in the sample.

COMPONENTS

ichroma™ Tn-I Plus consists of 'cartridges', 'detector tubes' and 'detector diluent'.

- The cartridge contains the membrane called a test strip which has streptavidin at the test line, and chicken IgY at the control line. All cartridges are individually sealed in an aluminum foil pouch containing a desiccant, and they are further packaged in a box.
- The detector tube has 2 granules containing anti-Tn-I-fluorescence conjugate, anti-chicken IgY-fluorescence conjugate, biotin-anti-Tn-I conjugate and sodium azide as a preservative in Tris-Cl buffer. All detector tubes are packed in a pouch.
- The detector diluent contains Tween 20 as a surfactant and sodium azide as a preservative in Tris-Cl buffer, and it is pre-dispensed in a vial. The detector diluent is packed in a box.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Follow the instructions and procedures described in this 'Instructions for use'.
- Use only fresh samples and avoid direct sunlight.
- Lot numbers of all the test components (cartridge, detector tube, detector diluent and ID chip) must match each other.
- Do not interchange the test components between different lots or use the test components after the expiration date, either of which might yield incorrect test result(s).
- Do not reuse cartridges, detector tubes or capillary tubes. A cartridge should be used for testing one sample only. A detector tube should be used for processing of one sample only. A capillary tube should be used for testing one sample only.
- The cartridge should remain sealed in its original pouch until just before use. Do not use cartridge, if pouch is damaged or has already been opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with local regulations. Sample with severe hemolysis and/or hyperlipidemia must not be used.
- If test components and/or sample are stored in refrigerator, then allow cartridge, detector tube, detector diluent and sample to be at room temperature for approximately 30 minutes before use.
- The instrument for ichroma™ tests may generate slight vibration during use.
- Used cartridges, detector tubes, detector diluent, capillary tube and pipette tips should be handled carefully and discarded by an appropriate method in accordance with relevant local regulations.
- The detector tube and the detector diluent contain sodium azide (NaN_3), and it may cause certain health issues like convulsions, low blood pressure, low heart rate, loss of consciousness, lung injury and respiratory failure. Avoid contact with skin, eyes, and clothing. In case of contact, rinse immediately with running water.
- No Biotin interference was observed in **ichroma™ Tn-I Plus** when biotin concentration in the sample was below 5.0 ng/mL. If a patient has been taking biotin at dosage of more than 0.03 mg a day, it is recommended to test again 24 hours after discontinuation of biotin intake.
- **ichroma™ Tn-I Plus** will provide accurate and reliable results subject to the below conditions.
 - **ichroma™ Tn-I Plus** should be used only in conjunction with the instrument for ichroma™ tests.
 - Have to use recommended anticoagulant.

Recommended anticoagulant
Sodium heparin, Lithium heparin, Sodium citrate
- **The capillary tube should be used when the following conditions are met.**
 - The capillary tube provided with the kit is recommended to obtain correct test result.
 - Whole blood should be immediately tested after collection.
 - Excess whole blood around the capillary tube should be wiped off.

- In order to avoid cross-contamination, please do not re-use capillary tube for multiple samples.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antibodies.
- The test may yield false negative result(s) due to the non-responsiveness of the antigens to the antibodies which is the most common if the epitope is masked by some unknown components, so therefore not being able to be detected or captured by the antibodies. The instability or degradation of the antigens with time and/or temperature may also cause false negative result as it makes the antigens unrecognizable by the antibodies.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician in conjunction with clinical symptoms and other relevant test results.

STORAGE AND STABILITY

Component	Storage condition		
	Storage Temperature	Shelf life	Note
Cartridge	2 – 30°C	20 months	Disposable
Detector tube	2 – 30°C	20 months	Disposable
Detector diluent	2 – 30°C	20 months	Unopened
		20 months	Opened

- After the cartridge pouch is opened, the test should be performed immediately.

MATERIALS SUPPLIED

REF CFPC-65

Components of **ichroma™ Tn-I Plus**

- Cartridge box:
 - Cartridge 25
 - Detector tube 25
 - 50 µL Capillary tube 25
 - Detector diluent 1
 - ID chip 1
 - Instructions for use 1

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately with **ichroma™ Tn-I Plus**.

Please contact our sales division for more information.

■ Instrument for **ichroma™ tests**

- **ichroma™ II** REF FPRR021
- **ichroma™ III** REF FPRR037
- **ichroma™ M2** REF FPRR031
- **ichroma™-50** REF FPRR022
- **ichroma™-50 PLUS** REF FPRR036
- **Boditech Tn-I Plus Control** REF CFPO-212

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ Tn-I Plus** is human whole blood/serum/plasma.

- It is recommended to test the sample within 24 hours after collection when collected sample is stored at room temperature.
- The samples (serum, plasma) should be separated from the clot by centrifugation within 3 hours after the collection of whole blood.
- The samples (whole blood, serum, plasma) may be stored for a week at 2 - 8 °C prior to being tested. If testing will be delayed more than a week, samples (serum, plasma) should be frozen at -20 °C.
- The samples (serum, plasma) stored frozen at -20 °C for 3 months showed no performance difference.
- However, the whole blood sample should not be kept in a freezer in any case.
- As a repeated freeze-thaw cycle may affect the test result, do not refreeze previously frozen samples.
- Whole blood sample can be collected using a capillary tube according to below:
 - ① Wear disposable gloves and protective equipment for safety.
 - ② Open the zipper bag which has capillary tubes.
 - ③ Take out the capillary tube and check for damage or contamination.
 - ④ Hold the handle of the capillary tube and touch the surface of blood with the capillary tube.
 - ⑤ Fill it with blood completely.
(Make sure that no air bubbles are present in the capillary tube. Do not get blood on the surface of the capillary tube. If the blood gets on the surface of the capillary tube, remove it gently with gauze.)

TEST SETUP

- Check the contents of **ichroma™ Tn-I Plus**: Sealed cartridges, detector tubes, a detector diluent, capillary tubes, an ID chip and an instructions for use.
- Ensure that the lot number of the cartridges matches that of the detector tubes, the detector diluent as well as the ID chip.
- If the sealed cartridges, the detector tube and the detector diluent have been stored in a refrigerator, place them on a clean and flat surface at room temperature for at least 30 minutes before testing.
- Turn on the instrument for **ichroma™ tests**.
- Insert the ID chip into the 'ID chip port'.

※ Please refer to the instrument for **ichroma™ tests** operation manual for complete information and operating instructions.

TEST PROCEDURE

▶ **ichroma™ II, ichroma™ M2**

Multi test mode/ Read now mode

- 1) Select the sample type (whole blood or serum/plasma) on the screen.
 - 2) Take 150 µL of detector diluent using a pipette and dispense it to the detector tube containing a granule. When the granule form is completely dissolved in the tube, it becomes detection buffer. (The detection buffer must be used immediately. Do not exceed 30 seconds.)
 - 3) Take 50 µL of sample (whole blood/serum/plasma/control) using a pipette and dispense it to the detector tube.
 - 4) Close the lid of the detector tube and mix the sample thoroughly by shaking it about 20 times. (The sample mixture must be used immediately. Do not exceed 30 seconds.)
 - 5) Take 75 µL of the sample mixture and dispense it into the sample well of the cartridge.
 - 6) Leave the cartridge at room temperature for 12 minutes.
- ⚠ Scan the sample-loaded cartridge immediately when the incubation time is over. If not, it will cause inaccurate test result.
- 7) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
 - 8) Tap the 'Start' button on the instrument for ichroma™ tests to start the scanning process. (ichroma™ M2 will start the test automatically after inserting.)
 - 9) The instrument for ichroma™ tests will start scanning the sample-loaded cartridge immediately.
 - 10) Read the test result on the display screen of the instrument for ichroma™ tests.

Single test mode/ Walk away mode

- 1) The test procedure is same with the 'Multi test mode 1) – 5)'.
- 2) Insert the sample-loaded cartridge into the holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 3) Tap the 'Start' button on the instrument for ichroma™ tests. (ichroma™ M2 will start the test automatically after inserting.)
- 4) The cartridge goes inside the instrument for ichroma™ tests and will automatically start scanning the sample-loaded cartridge after 12 minutes.
- 5) Read the test result on the display screen of the instrument for ichroma™ tests.

▶ **ichroma™ III**

- 1) The test procedure is same with the 'Single test mode'.

▶ **ichroma™-50, ichroma™-50 PLUS**

- 1) Insert the tip array in the tip station.
- 2) Insert the detector tube in the reagent station and cover the reagent station to hold the detector tubes in place.
- 3) Open the lid of the detector diluent and insert the detector diluent in the diluent station.
- 4) Insert the cartridge magazine with the cartridges into the magazine station.
- 5) Insert the sample tube into the blood collection tube rack and load the blood collection tube rack into the sampling station (loading part).
- 6) Tap the button located in the upper side of the No. of test cartridge region to select the ID chip that you want to use.
- 7) When the selected cartridge slot is activated, set the number of the detector tube by tapping.
- 8) Set the number of pipette tips by tapping.
- 9) Select the sample type (whole blood or serum/plasma) on the screen.
- 10) Tap the 'Start' button on the left upper of the main screen to start test.

INTERPRETATION OF TEST RESULT

- The instrument for ichroma™ tests calculates the test result automatically and displays Tn-I concentration of the test sample in terms of ng/mL.
- Working range: 0.01-15.00 ng/mL.
- **Expected Values**
 - In studies performed with the **ichroma™ Tn-I Plus** assay involving 125 healthy volunteers in Korea, the upper reference limit (99th percentile) for Tn-I was 0.04 ng/mL. The lowest concentration with a CV less than or equal to 10 % with the **ichroma™ Tn-I Plus** assay was 0.04 ng/mL.
 - Due to the release kinetics of Tn-I, a result below the decision limit within the first hours of the onset of symptoms does not rule out myocardial infarction with certainty. If myocardial infarction is still suspected, repeat the test at appropriate intervals.

QUALITY CONTROL

- Quality control tests are a part of the good testing practice to confirm the expected results and validity of the assay and should be performed at regular intervals.
- Quality control tests should also be performed whenever there is any question concerning the validity of the test results.
- Control materials are not provided on demand with **ichroma™ Tn-I Plus**. For more information regarding obtaining the control materials, contact **Boditech Med Inc.'s Sales Division for assistance**. (Please refer to the instructions for use of control material.)

PERFORMANCE CHARACTERISTICS

Analytical sensitivity

- Limit of Blank (LoB) 0.004 ng/mL
- Limit of Detection (LoD) 0.01 ng/mL
- Limit of Quantitation (LoQ) 0.03 ng/mL

Analytical specificity

- Cross-reactivity

Biomolecules listed in the following table were added to the test sample(s) at concentrations much higher than their normal physiological levels in the blood. **ichroma™ Tn-I Plus** test results did not show any significant cross-reactivity with these biomolecules.

Cross-reactants	Concentration
CK-MB	60 ng/mL
NT-proBNP	1,000 ng/mL
Myoglobin	1,000 ng/mL
D-Dimer	1,000 ng/mL
Myosin	1,000 ng/mL
Cardiac troponin C	250 ng/mL
Skeletal troponin I	250 ng/mL
Tropomyosin	1,000 ng/mL
Cardiac troponin T	125 ng/mL
Actin	1,000 ng/mL

- Interference

Interferents listed in the following table were added to the test sample at the concentration mentioned below. **ichroma™ Tn-I Plus** test results did not show any significant interference with these materials except for EDTA.

Interferents	Concentration
Bilirubin	350 µmol/L
Cholesterol	13 mmol/L
D-Glucose	1,000 mg/dL
Hemoglobin	2 g/L
L-Ascorbic acid	350 µmol/L
Triglyceride mixture	500 mg/dL
Sodium Heparin	3,000 U/L
Li Heparin	3,000 U/L
Sodium citrate	2 mg/mL
EDTA-K2	3.4 µmol/L
EDTA-K3	3.4 µmol/L

Precision

- Single-site study

Repeatability (within-run precision)

within-laboratory precision (Total precision)

Lot to lot precision

3 Lots of **ichroma™ Tn-I Plus** were tested for 20 days. Each standard material was tested 2 times per day. For each test, each material was duplicated.

Tn-I Plus [ng/mL]	Single-site study					
	Repeatability		within-laboratory precision		Lot to lot precision	
	AVG [ng/mL]	CV (%)	AVG [ng/mL]	CV (%)	AVG [ng/mL]	CV (%)
0.23	0.23	7.0	0.24	7.1	0.23	7.2
0.94	0.95	6.2	0.95	7.4	0.95	6.9
7.50	7.55	7.4	7.55	8.2	7.57	7.6

Multi-site study

Reproducibility

1 Lot of **ichroma™ Tn-I Plus** was tested for 5 days in 3 different sites (1 person per 1 site, 1 instrument per 1 site). Each standard material was tested 1 time per and 5 replicates per day.

Tn-I Plus [ng/mL]	Multi-site study	
	Reproducibility	
	AVG [ng/mL]	CV (%)
0.23	0.23	7.7
0.94	0.95	7.0
7.50	7.67	7.2

Accuracy

The accuracy was confirmed by testing with 3 different lots of **ichroma™ Tn-I Plus**. The tests were repeated 10 times at each concentration of the control standard.


Tn-I Plus [ng/mL]	Lot 1	Lot 2	Lot 3	AVG [ng/mL]	Recovery (%)
0.05	0.05	0.05	0.05	0.05	96.7
0.62	0.61	0.58	0.58	0.59	94.8
1.52	1.40	1.48	1.50	1.46	96.0
2.42	2.37	2.34	2.43	2.38	98.3
3.02	2.86	2.96	2.78	2.86	94.8
4.01	3.83	3.95	3.92	3.90	97.3
6.01	5.63	5.79	5.74	5.72	95.2
12	11.91	11.76	11.71	11.79	98.3

Comparability

Tn-I concentrations of 100 clinical samples were quantified independently with **ichroma™ Tn-I Plus (ichroma™ II)** and **Comparator A** as per prescribed test procedures. Test results were compared, and their comparability was investigated with linear regression and correlation coefficient (R). The regression equation and correlation coefficient are as follows.

* Solid line: regression line of **ichroma™ Tn-I Plus**













* Dash line: $y = x$



REFERENCES

1. Mauro Panteghini, Franca Pacani, Kiang-Teck J.Yeo, Fred S. Apple, Robert H. Christenson, Francesco Dati, Johannes Mair, Jan Ravkilde, and Alan H.B. We. Evaluation of Imprecision for Cardiac Troponin Assays at Low-Range Concentrations. 2004;50:2:327-332.
2. Alan McNeil, PhD, FRACP, FRCPA. The Trouble with Troponin. Heart, Lung and Circulation 2007;16:S13-S16.
3. David M. Bunk and Micahel J. Welch. Characterization of a New Certified Reference Material for Human Cardiac Troponin I. Clinical Chemistry 2002;52:2:212-219
4. Jaffe AS, Ravkilde J, Roberts R, Naslund U, Apple FS, Galvani M, Katus H. It's time for a change to a troponin standard. Circulation 2000;102:1216 –1220.
5. Jillan R. Tate, David Heathcote, Gus Koerbin, Gary Thean, David Andriske, Jone Bonar, Janice Gill. The harmonization of cardiac troponin I measurement is independent of sample time collection but is dependent on the source of calibrator. Clinica Chimica Acta 324:2002:13-23.
6. Ohman EM, Armstrong PW, Christenson RH, et al. Cardiac troponin T levels for risk stratification in acute myocardial ischemia. N Engl J Med 1996;335:1333– 41.
7. Antman EM, Tanasijevic MJ, Thompson B, et al. Cardiac-specific troponin I levels to predict the risk of mortality in patients with acute coronary syndromes. N Engl J Med 1996;335:1342 –9.

Note: Please refer to the table below to identify various symbols

	Sufficient for <n> tests
	Read instruction for use
	Use by Date
	Batch code
	Catalog number
	Caution
	Manufacturer
	Authorized representative of the European Community
	In vitro diagnostic medical device
	Temperature limit
	Do not reuse
	This product fulfills the requirements of the Directive 98/79/EC on in vitro diagnostic medical devices

For technical assistance, please contact:

Boditech Med Inc.'s Technical Sales

Tel: +(82) -33-243-1400

E-mail: TS@boditech.co.kr



Boditech Med Inc.

43, Geodudanji 1-gil, Dongnae-myeon, Chuncheon-si, Gang-won-do, 24398, Republic of Korea

Tel: +(82) -33-243-1400

Fax: +(82) -33-243-9373

www.boditech.co.kr



Obelis s.a

Bd. Général Wahis 53, 1030 Brussels, Belgium

Tel: +(32) -2-732-59-54

Fax: +(32) -2-732-60-03

E-Mail: mail@obelis.net

