



0123



130252007M: 100 tests/kit
130652007M: 50 tests/kit
130752007M: 30 tests/kit

MAGLUMI® Estradiol (CLIA)

INTENDED USE

The kit is an *in vitro* chemiluminescence immunoassay for the quantitative determination of Estradiol (E2) in human serum and plasma using the MAGLUMI series Fully-auto chemiluminescence immunoassay analyzer and Biolumi series Integrated System, and the assay is used for an aid in the diagnosis and treatment of individuals with suspected or confirmed various hormonal sexual disorders.

SUMMARY

Estrogens are steroid hormones primarily known for their role in promotion of female sex characteristics and reproductive capability. There are three forms of estrogens in the female body: estrone (E1), estradiol (E2), and estriol (E3). During a woman's reproductive years, the principal circulating estrogen is 17 β -estradiol (E2); importantly, it is also the most potent form of estrogen¹. In females, estradiol (E2) is produced mainly by the developing follicles in the ovary, the corpus luteum and the placenta, while in males, E2 is produced in the testes. Small amounts of E2 are also produced by the liver, adrenal glands, and breasts. In postmenopausal women, estrone is the main circulating estrogen synthesized from dehydroepiandrosterone and secreted by the adrenals². Estradiol (E2) in serum serves as an important diagnostic marker in a variety of clinical conditions in both men and women. Clinically, serum levels of E2 are used to assess ovarian function in women with menstrual disorders, precocious or delayed puberty, and assisted reproduction³, hirsutism, polycystic ovary syndrome (PCOS)⁴. Monitoring of E2 levels across pregnancy can be of use in pregnancy-related disorders. E2 levels together with progesterone and inhibin B levels predict the risk of hydatidiform mole in patients with molar pregnancies⁵. Monitoring anti-estrogen or aromatase inhibitor treatment for pubertal delay (with or without short stature), obesity, or male infertility and a threshold of circulating E2 for maintenance of bone density in older men⁶.

TEST PRINCIPLE

Sandwich chemiluminescence immunoassay.

The sample, magnetic microbeads coated with anti-E2 antibody and buffer are mixed thoroughly, incubating and performing a wash cycle after a precipitation in a magnetic field. ABEI labeled with anti-E2 antibody are then added, reacting to form sandwich complexes and incubating. After precipitation in a magnetic field, the supernatant is decanted and then a wash cycle is performed. Subsequently, the Starter 1+2 are added to initiate a chemiluminescent reaction. The light signal is measured by a photomultiplier as relative light units (RLUs), which is proportional to the concentration of E2 present in the sample.

REAGENTS

Kit Contents

Component	Description	100 tests/kit	50 tests/kit	30 tests/kit
Magnetic Microbeads	Magnetic microbeads coated with anti-E2 antibody (~10.0 μ g/mL) in PBS buffer, NaNa ₃ (<0.1%).	2.5 mL	1.5 mL	1.0 mL
Calibrator Low	A low concentration of E2 antigen, BSA, NaNa ₃ (<0.1%).	1.0 mL	1.0 mL	1.0 mL
Calibrator High	A high concentration of E2 antigen, BSA, NaNa ₃ (<0.1%).	1.0 mL	1.0 mL	1.0 mL
Buffer	ANS, Tris-HCl buffer, NaNa ₃ (<0.1%).	10.5 mL	6.0 mL	4.2 mL
ABEI Label	ABEI labeled with anti-E2 antibody (~0.500 μ g/mL) in Tris-HCl buffer, NaNa ₃ (<0.1%).	17.5 mL	9.5 mL	6.3 mL
Diluent	BSA, NaNa ₃ (<0.1%).	5.5 mL	3.5 mL	3.5 mL
Control 1	A low concentration of E2 antigen (100 pg/mL), BSA, NaNa ₃ (<0.1%).	1.0 mL	1.0 mL	1.0 mL
Control 2	A high concentration of E2 antigen (400 pg/mL), BSA, NaNa ₃ (<0.1%).	1.0 mL	1.0 mL	1.0 mL
All reagents are provided ready-to-use.				

The control barcode labels are provided.

Warnings and Precautions

- For *in vitro* diagnostic use.
- For professional use only.
- Exercise the normal precautions required for handling all laboratory reagents.
- Personal protective measures should be taken to prevent any part of the human body from contacting samples, reagents, and controls, and should comply with local operating requirements for the assay.
- A skillful technique and strict adherence to the package insert are necessary to obtain reliable results.
- Do not use kit beyond the expiration date indicated on the label.
- Do not interchange reagent components from different reagents or lots.
- Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).
- All waste associated with biological samples, biological reagents and disposable materials used for the assay should be considered potentially infectious and should be disposed of in accordance with local guidelines.
- This product contains sodium azide. Sodium azide may react with lead or copper plumbing to form highly explosive metal azides. Immediately after disposal, flush with a large volume of water to prevent azide build-up. For additional information, see Safety Data Sheets available for professional user on request.

Note: If any serious incident has occurred in relation to the device, please report to Shenzhen New Industries Biomedical Engineering Co., Ltd. (Snibe) or our authorized representative and the competent authority of the Member State in which you are established.

Reagent Handling

- To avoid contamination, wear clean gloves when operating with a reagent kit and sample. When handling reagent kit, replace the gloves that have been in contact with samples, since introduction of samples will result in unreliable results.
- Do not use kit in malfunction conditions; e.g., the kit leaking at the sealing film or elsewhere, obviously turbid or precipitation is found in reagents (except for Magnetic Microbeads) or control value is out of the specified range repeatedly. When kit in malfunction conditions, please contact Snibe or our authorized distributor.
- To avoid evaporation of the liquid in the opened reagent kits in refrigerator, it is recommended that the opened reagent kits to be sealed with reagent seals contained within the packaging. The reagent seals are single use, and if more seals are needed, please contact Snibe or our authorized distributor.
- Over time, residual liquids may dry on the septum surface. These are typically dried salts and have no effect on assay efficacy.
- Use always the same analyzer for an opened reagent integral.
- For magnetic microbeads mixing instructions, refer to the Preparation of the Reagent section of this package insert.
- For further information about the reagent handling during system operation, please refer to Analyzer Operating Instructions.

Storage and Stability

- Do not freeze the integral reagents.
- Store the reagent kit upright to ensure complete availability of the magnetic microbeads.
- Protect from direct sunlight.

Stability of the Reagents	
Unopened at 2-8°C	until the stated expiration date
Opened at 2-8°C	6 weeks
On-board	4 weeks

Stability of Controls	
Unopened at 2-8°C	until the stated expiration date
Opened at 10-30°C	6 hours
Opened at 2-8°C	6 weeks
Frozen at -20°C	3 months
Frozen and thawed cycles	no more than 3 times

SPECIMEN COLLECTION AND PREPARATION

Specimen Types

Only the specimens listed below were tested and found acceptable.

Specimen Types	Collection Tubes
Serum	Tubes without additive/accessory, or tubes containing clot activator or clot activator with gel.
Plasma	K2-EDTA

- The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. Follow tube manufacturers' instructions carefully when using collection tubes.

Specimen Conditions

- Do not use grossly hemolyzed/hyperlipidaemia specimens and specimens with obvious microbial contamination.
- Ensure that complete clot formation in serum specimens has taken place prior to centrifugation. Some serum specimens, especially those from patients receiving anticoagulant or thrombolytic therapy, may exhibit increased clotting time. If the serum specimen is centrifuged before a complete clotting, the presence of fibrin may cause erroneous results.
- Samples must be free of fibrin and other particulate matter.
- To prevent cross contamination, use of disposable pipettes or pipette tips is recommended.

Preparation for Analysis

- Inspect all specimens for foam. Remove foam with an applicator stick before analysis. Use a new applicator stick for each specimen to prevent cross contamination.
- Frozen specimens must be completely thawed before mixing. Mix thawed specimens thoroughly by low speed vortexing or by gently inverting. Visually inspect the specimens. If layering or stratification is observed, mix until specimens are visibly homogeneous. If specimens are not mixed thoroughly, inconsistent results may be obtained.
- Specimens should be free of fibrin, red blood cells, or other particulate matter. Such specimens may give reliable results and must be centrifuged prior to testing. Transfer clarified specimen to a sample cup or secondary tube for testing. For centrifuged specimens with a lipid layer, transfer only the clarified specimen and not the lipemic material.
- The sample volume required for a single determination of this assay is 20 μ L.

Specimen Storage

Specimens removed from the separator, red blood cells or clot may be stored up to 8 hours at 10-30°C, or 7 days at 2-8°C, or 6 months frozen at -20°C. Frozen specimens subjected to up to 2 freeze/thaw cycles have been evaluated.

Specimen Shipping

- Package and label specimens in compliance with applicable local regulations covering the transport of clinical specimens and infectious substances.
- Do not exceed the storage limitations listed above.

Specimen Dilution

- Samples, E2 concentrations above the analytical measuring interval, can be diluted with Diluent either automated dilution protocol or manual dilution procedure. The recommended dilution ratio is 1:10. The concentration of the diluted sample must be >480 pg/mL.
- For manual dilution, multiply the result by the dilution factor. For dilution by the analyzers, the analyzer software automatically takes the dilution into account when calculating the sample concentration.

PROCEDURE

Materials Provided

Estradiol (CLIA) assay, control barcode labels.

Materials Required (But Not Provided)

- General laboratory equipment.
- Fully-auto chemiluminescence immunoassay analyzer Maglumi 600, Maglumi 800, Maglumi 1000, Maglumi 2000, Maglumi 2000 Plus, Maglumi 4000, Maglumi 4000 Plus, MAGLUMI X8, MAGLUMI X3, MAGLUMI X6, or Integrated System Biolumi 8000 and Biolumi CX8.
- Additional accessories of test required for the above analyzers include Reaction Module, Starter 1+2, Wash Concentrate, Light Check, Tip, and Reaction Cup. Specific accessories and accessories' specification for each model refer to corresponding Analyzer Operating Instructions.
- Please use accessories specified by Snibe to ensure the reliability of the test results.

Assay Procedure

Preparation of the Reagent

- Take the reagent kit out of the box and visually inspect the integral vials for leaking at the sealing film or elsewhere. If there is no leakage, please tear off the sealing film carefully.
- Open the reagent area door; hold the reagent handle to get the RFID label close to the RFID reader (for about 2s); the buzzer will beep; one beep sound indicates successful sensing.
- Keeping the reagent straight insert to the bottom along the blank reagent track.
- Observe whether the reagent information is displayed successfully in the software interface, otherwise repeat the above two steps.
- Resuspension of the magnetic microbeads takes place automatically when the kit is loaded successfully, ensuring the magnetic microbeads are totally resuspended homogenous prior to use.

Assay Calibration

- Select the assay to be calibrated and execute calibration operation in reagent area interface. For specific information on ordering calibrations, refer to the calibration section of Analyzer Operating Instructions.
- Execute recalibration according to the calibration interval required in this package insert.

Quality Control

- When new lot used, check or edit the quality control information.
- Scan the control barcode, choose corresponding quality control information and execute testing. For specific information on ordering quality controls, refer to the quality control section of the Analyzer Operating Instructions.

Sample Testing

- After successfully loading the sample, select the sample in interface and edit the assay for the sample to be tested and execute testing. For specific information on ordering patient specimens, refer to the sample ordering section of the Analyzer Operating Instructions.

To ensure proper test performance, strictly adhere to Analyzer Operating Instructions.

Calibration

Traceability: This method has been standardized against the USP reference standard (Catalog number: 1250008).

Test of assay specific calibrators allows the detected relative light unit (RLU) values to adjust the master curve.

Recalibration is recommended as follows:

- Whenever a new lot of Reagent or Starter 1+2 is used.
- Every 28 days.

- The analyzer has been serviced.
- Control values lie outside the specified range.

Quality Control

Controls are recommended for the determination of quality control requirements for this assay and should be run in singlicate to monitor the assay performance. Refer to published guidelines for general quality control recommendations, for example Clinical and Laboratory Standards Institute (CLSI) Guideline C24 or other published guidelines.

Quality control is recommended once per day of use, or in accordance with local regulations or accreditation requirements and your laboratory's quality control procedures, quality control could be performed by running the E2 assay:

- Whenever the kit is calibrated.
- Whenever a new lot of Starter 1+2 or Wash Concentrate is used.

Controls are only applicable with MAGLUMI and Biolumi system and only used matching with the same top seven LOT numbers of corresponding reagents. For each target value and range refer to the label.

The performance of other controls should be evaluated for compatibility with this assay before they are used. Appropriate value ranges should be established for all quality control materials used.

Control values must lie within the specified range, whenever one of the controls lies outside the specified range, calibration should be repeated and controls retested. If control values lie repeatedly outside the predefined ranges after successful calibration, patient results must not be reported and take the following actions:

- Verify that the materials are not expired.
- Verify that required maintenance was performed.
- Verify that the assay was performed according to the package insert.
- If necessary, contact Snibe or our authorized distributors for assistance.

If the controls in kit are not enough for use, please order Estradiol (CLIA) Controls (REF: 160201256MT) from Snibe or our authorized distributors for more.

RESULTS

Calculation

- The analyzer automatically calculates the E2 concentration in each sample by means of a calibration curve which is generated by a 2-point calibration master curve procedure. The results are expressed in pg/mL. For further information please refer to the Analyzer Operating Instructions.

- Conversion factors: pmol/L x 0.272 = pg/mL(ng/L);
pg/mL x 3.67 = pmol/L

Interpretation of Results

The expected range for the E2 assay was obtained by testing 1059 apparently healthy individuals in China, gave the following expected value:

Test subjects	N	Mean (pg/mL)	2.5 th percentile (pg/mL)	5 th percentile (pg/mL)	95 th percentile (pg/mL)	97.5 th percentile (pg/mL)
Males	150	38.265	24.2	/	/	62.1
Non-pregnant Females	Follicular Phase	130	71.258	11.2	/	225
	Ovulation Phase	127	161.531	36.0	/	403
	Luteal Phase	142	125.268	21.0	/	343
	Postmenopause	134	38.593	/	136	/
	1st Trimester	126	1078.658	152	/	3120
Pregnancy	2nd Trimester	127	8733.784	1423	/	22032
	3rd Trimester	123	21807.685	/	8413	/

Results may differ between laboratories due to variations in population and test method. It is recommended that each laboratory establish its own reference interval.

LIMITATIONS

- Results should be used in conjunction with patient's medical history, clinical examination and other findings.
- If the E2 results are inconsistent with clinical evidence, additional testing is needed to confirm the result.
- Heterophilic antibodies in human serum can react with reagent immunoglobulins, interfering with *in vitro* immunoassays. Patients routinely exposed to animals or animal serum products can be prone to this interference and anomalous values may be observed⁸.
- Bacterial contamination of the specimens may affect the test results.

SPECIFIC PERFORMANCE CHARACTERISTICS

Representative performance data are provided in this section. Results obtained in individual laboratories may vary.

Precision

Precision was determined using the assay, samples and controls in a protocol (EP05-A3) of the CLSI (Clinical and Laboratory Standards Institute): duplicates at two independent runs per day for 5 days at three different sites using three lots of reagent kits (n = 180). The following results were obtained:

Sample	Mean (pg/mL) (n=180)	Within-Run		Between-Run		Reproducibility	
		SD (pg/mL)	%CV	SD (pg/mL)	%CV	SD (pg/mL)	%CV
Serum Pool 1	34.493	1.399	4.06	0.655	1.90	2.219	6.43
Serum Pool 2	201.379	6.733	3.34	2.442	1.21	9.379	4.66
Serum Pool 3	405.437	13.660	3.37	5.747	1.42	20.486	5.05
Plasma Pool 1	35.028	1.290	3.68	0.539	1.54	1.583	4.52
Plasma Pool 2	203.255	5.839	2.87	5.152	2.53	9.639	4.74
Plasma Pool 3	405.326	13.727	3.39	9.584	2.36	18.322	4.52
Control 1	101.048	3.914	3.87	2.789	2.76	6.248	6.18
Control 2	390.597	13.832	3.54	7.223	1.85	20.361	5.21

Linear Range

10.0-4800 pg/mL (defined by the Limit of Quantitation and the maximum of the master curve).

Reportable Interval

5.00-48000 pg/mL (defined by the Limit of Detection and the maximum of the master curve*Recommended Dilution Ratio).

Analytical Sensitivity

Limit of Blank (LoB) =1.00 pg/mL.

Limit of Detection (LoD) =5.00 pg/mL.

Limit of Quantitation (LoQ) =10.0 pg/mL.

Analytical Specificity

Interference

Interference was determined using the assay, three samples containing different concentrations of analyte were spiked with potential endogenous and exogenous interferents in a protocol (EP7-A2) of the CLSI. The measurement deviation of the interference substance is within ±10%. The following results were obtained:

Interference	No interference up to	Interference	No interference up to
Bilirubin	20 mg/dL	Rheumatoid factor	1500 IU/mL
Hemoglobin	1000 mg/dL	ANA	398 AU/mL
Intralipid	2000 mg/dL	Biotin	0.5 mg/dL
Estrone sulfate	1000 ng/mL	Oestrone	20000 pg/mL
Estriol 3-sulfate	2500 ng/mL	Estriol 17-sulfate	2500 ng/mL
Ethinylestradiol	50000 pg/mL	Estradiol Valerate	1000 ng/mL

Aldosterone	10000 ng/mL	17β-estradiol 3 glucosidate	100000 pg/mL
Estradiol 3,17β diglucosidate	2000 ng/mL	Ethisterone	1000 ng/mL
Testosterone	10000 ng/mL	Estrone-3β-glucosidate	1000 ng/mL

Cross-Reactivity

Cross-reactivity was determined using the assay, three samples containing different concentrations of analyte were spiked with potential cross-reactants in a protocol (EP7-A2) of the CLSI. The measurement deviation of the interference substance is within ±10%. The following results were obtained:

Cross-reactant	No interference up to	Cross-reactant	No interference up to
Estriol	10000 pg/mL	17 –α-estradiol	100000 pg/mL
Androstenediol	2000 ng/mL		

High-Dose Hook

No high-dose hook effect was seen for E2 concentrations up to 100000 pg/mL.

Method Comparison

A comparison of the E2 assay with a commercially available immunoassay, gave the following correlations (pg/mL):

Number of samples measured: 145

Passing-Bablok: y=0.9940x+0.3029, r=0.982.

The clinical specimen concentrations were between 7.53 and 4757 pg/mL.

REFERENCES

1. Rettberg J R, Yao J, Brinton R D. Estrogen: a master regulator of bioenergetic systems in the brain and body[J]. Frontiers in neuroendocrinology, 2014, 35(1): 8-30.
2. Kumar A, Banerjee A, Singh D, et al. Estradiol: A steroid with multiple facets[J]. Hormone and Metabolic Research, 2018, 50(05): 359-374.
3. Stanczyk F Z, Jurov J, Hsing A W. Limitations of direct immunoassays for measuring circulating estradiol levels in postmenopausal women and men in epidemiologic studies[J]. Cancer Epidemiology and Prevention Biomarkers, 2010, 19(4): 903-906.
4. Ketha H, Girtman A, Singh R J. Estradiol assays—the path ahead[J]. Steroids, 2015, 99: 39-44.
5. Rosner W, Hankinson S E, Sluss P M, et al. Challenges to the measurement of estradiol: an endocrine society position statement[J]. The Journal of Clinical Endocrinology & Metabolism, 2013, 98(4): 1376-1387.
6. Handelsman D J, Newman J D, Jimenez M, et al. Performance of direct estradiol immunoassays with human male serum samples[J]. Clinical Chemistry, 2014, 60(3): 510-517.
7. CLSI. Statistical Quality Control for Quantitative Measurement Procedures: Principles and Definitions. 4th ed. CLSI guideline C24. Wayne, PA: Clinical and Laboratory Standards Institute; 2016.
8. Boscato L M, Stuart M C. Heterophilic antibodies: a problem for all immunoassays [J]. Clinical Chemistry, 1988, 34(1):27-33.

SYMBOLS EXPLANATIONS

	Consult instructions for use		Manufacturer
	Temperature limit (Store at 2-8 °C)		Use-by date
	Contains sufficient for<n> tests		Keep away from sunlight
	This way up		Authorized representative in the European Community
	In vitro diagnostic medical device		Kit component
	Catalogue number		Batch code
	CE marking with notified body ID number		

MAGLUMI® and Biolumi® are trademarks of Snibe. All other product names and trademarks are the property of their respective owners.

	Shenzhen New Industries Biomedical Engineering Co., Ltd.		
	No.23, Jinxiu East Road, Pingshan District, 518122 Shenzhen, P.R. China		
	Tel: +86-755-21536601	Fax: +86-755-28292740	
	Shanghai International Holding Corp. GmbH (Europe)		
	Eiffestrasse 80, 20537 Hamburg, Germany		
	Tel: +49-40-2513175		Fax: +49-40-255726